

RN-Regular Neck Connections

RN Abutment Screw Torque Instructions



1) The RN **Titanium** custom abutment is tightened into the implant to 35Ncm.

1) The 1.5 synOcta abutment is tightened into the implant to 35Ncm.

2) The RN **Zirconia** custom abutment is tightened into the 1.5 synOcta abutment to 15Ncm.

RN - ZrO₂ requires use of 1.5 synOcta

